

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | Fax: +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## HPV (Type 1, 6, 11, 16, 18, 31) Human Papilloma virus (Screening Antibody), Clone: [1H8], Mouse, Monoclonal NMB-HP300

| Article Name               | HPV (Type 1, 6, 11, 16, 18, 31) Human Papilloma virus (Screening Antibody), Clone: [1H8], Mouse, Monoclonal   |
|----------------------------|---|
| Biozol Catalog Number      | NMB-HP300   |
| Supplier Catalog Number    | HP300   |
| Alternative Catalog Number | NMB-HP300   |
| Manufacturer               | NordicMubio   |
| Host                       | Mouse   |
| Category                   | Antikörper  |
| Application                | IHC-Fr, IHC-P   |
| Immunogen                  | SDS-disrupted bovine papillomavirus type 1  |
| Product Description        | This antibody has been used for screening of HPV infections in various tissues. If positive staining turnes up, the analysis should be followed by testing for HPV 16 and 18 reactivity using antibody HP305 and/or specific PCR analysis. Human Papilloma          |
| Clonality                  | Monoclonal  |
| Clone Designation          | [1H8]   |
| Isotype                    | IgG   |
| Buffer                     | Antibody solution in stabilizing phosphate buffer pH 7.3. Contains 0.09 % sodium azide**. The volume is sufficient for at least 30 immunohistochemical tests (100 $\mu$ l working solution / test). Use appropriate antibody diluent e.g. BIOLOGO Art. No. PU002 if |

| Purity            | Antibody solution in stabilizing phosphate buffer pH 7.3. Contains 0.09 % sodium azide**. The volume is sufficient for at least 30 immunohistochemical tests (100 $\mu$ l working solution / test). Use appropriate antibody diluent e.g. BIOLOGO Art. No. PU002 if |
|-------------------|---|
| Application Notes | IHC(C, P)   |