

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Goat Antiserum anti-Human Secretory Component, free and boundunconj., MinX none NMB-GAHU/SC

Article Name	Goat Antiserum anti-Human Secretory Component, free and bound-unconj., MinX none
Biozol Catalog Number	NMB-GAHU/SC
Supplier Catalog Number	GAHu/SC
Alternative Catalog Number	NMB-GAHU/SC
Manufacturer	NordicMubio
Host	Goat
Category	Antikörper
Species Reactivity	Human
Conjugation	Unconjugated
Format	Antiserum
Target Specificity	Secretory Component - free and bound determinants
Cross-Adsorption (MinX)	no cross-adsorbtion
Product Description	Tested in immunoelectrophoresis, double radial immunodiffusion and ELISA against a panel of appropriate secretions and purified Ig isotypes. The antiserum reacts with both bound secretory component (secretory IgA) and with the free SC present in huma
Clonality	Polyclonal
Clone Designation	[Polyclonal]

Buffer	Delipidated, heat inactivated, lyophilized, stable whole antiserum No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of pooled normal goat serum. No foreign proteins added. Reconstitute the lyophilized a
Source	Secretory component is present in human secretions bound to secretory IgA (slgA) and in free form. Secretory IgA (slgA) functions as a dimer or polymer and accounts for almost all specific mucosal antibody activity. A molecule of slgA is made up of two m
Formula	Delipidated, heat inactivated, lyophilized, stable whole antiserum No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of pooled normal goat serum. No foreign proteins added.
Antibody Type	Secondary Antibody
Application Notes	Precipitation assays. In immunoelectrophoresis use 2 μ l serum or equivalent against 120 μ l antiserum. In double radial immunodiffusion use a rosette arrangement with 10 μ l antiserum in 3 mm diameter centre well and 2 μ l serum samples (neat and serially d