



Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## CLIA Kit for Substance P (SP) USC-CCA393BO

Article Name	CLIA Kit for Substance P (SP)
Biozol Catalog Number	USC-CCA393BO
Supplier Catalog Number	CCA393Bo
Alternative Catalog Number	USC-CCA393BO-48,USC-CCA393BO-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Bovine
Range	11.7-3,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 4.5pg/mL
UniProt	P01289
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.