



Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Substance P (SP) USC-CCA393GU

Article Name	CLIA Kit for Substance P (SP)
Biozol Catalog Number	USC-CCA393GU
Supplier Catalog Number	CCA393Gu
Alternative Catalog Number	USC-CCA393GU-48,USC-CCA393GU-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Guinea pig
Range	11.72-3,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 5.12pg/mL
UniProt	P67932
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.