

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | Fax: +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for 8-Hydroxydeoxyguanosine (8-OHdG) USC-CCA660GE

Article Name	CLIA Kit for 8-Hydroxydeoxyguanosine (8-OHdG)
Biozol Catalog Number	USC-CCA660GE
Supplier Catalog Number	CCA660Ge
Alternative Catalog Number	USC-CCA660GE-48,USC-CCA660GE-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	AII
Range	23.44-6,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 7.97pg/mL
Samples	serum, plasma and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.