

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human CD28(Cluster Of Differentiation 28) ELISA Kit EBT-ELK0672

Article Name	Human CD28(Cluster Of Differentiation 28) ELISA Kit
Biozol Catalog Number	EBT-ELK0672
Supplier Catalog Number	ELK0672
Alternative Catalog Number	EBT-ELK0672-96,EBT-ELK0672-48,EBT-ELK0672-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	50 ng/mL
Range	0.78-50 ng/mL
Sensitivity	0.32 ng/mL
UniProt	P10747
Samples	serum, plasma, cell lysates and other biological fluids

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule,Tumor immunity,Infection immunity,Immune molecule,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CD28. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CD28. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CD28, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CD28 in the samples is then determined by comparing the OD of the samples to the standard curve