

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **CircuLex S100A7/Psoriasis ELISA Kit MBL-CY-8073**

Article Name	CircuLex S100A7/Psoriasis ELISA Kit
Biozol Catalog Number	MBL-CY-8073
Supplier Catalog Number	CY-8073
Alternative Catalog Number	MBL-CY-8073
Manufacturer	MBL
Category	Kits/Assays
Application	ELISA
Product Description	CircuLex S100A7/Psoriasis ELISA Kit...
Range	Dilution factors need to be taken into consideration in calculating the human S100A7/Psoriasis concentration. Results exceeding human S100A7/Psoriasis level of 90 ng/ml should be repeated with diluted samples.
Sensitivity	better than 0.12 ng/ml of sample.
NCBI	<a href="#">6278</a>
Target	S100A7/Psoriasis