



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Estrone ELISA BVD-RCD010R

Article Name	Estrone ELISA
Biozol Catalog Number	BVD-RCD010R
Supplier Catalog Number	RCD010R
Alternative Catalog Number	BVD-RCD010R-96
Manufacturer	BioVendor
Category	Kits/Assays
Application	Serum
Species Reactivity	Human
Product Description	Competitive ELISA, Immobilized antigen
Range	10-2000 pg/ml
Sensitivity	3 pg/ml
Tag	BioVendor
Storage	Store the complete kit at 2–8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).
Target	Estrone
Application Notes	Intra assay: $CV = 8.2\%$ . Inter assay: $CV = 8.9\%$ . Spiking recovery: 101,9%. Dilution linearity: 95,50%