

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | Fax: +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## **GENLISA Human Cluster of Differentiation 89 (CD89 / FCAR) ELISA KBS-KB3077**

Article Name	GENLISA Human Cluster of Differentiation 89 (CD89 / FCAR) ELISA
Biozol Catalog Number	KBS-KB3077
Supplier Catalog Number	KB3077
Alternative Catalog Number	KBS-KB3077-1
Manufacturer	Krishgen Biosystems
Category	Kits/Assays
Species Reactivity	Human
Product Description	The GENLISA Human CD89 ELISA includes features like: - Ready to use protocol with break-apart wells for ease of use - Standardisation and High Reproducibility - Lot to Lot Consistency - Accuracy and Precision Validated against seven points for a "GO
Range	7.81 - 500 pg/ml
Sensitivity	3.5 pg/ml
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids. For tear samples, please connect with us to discuss further.
Application Notes	Colorimetric, Absorbance measured at 450nm