

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Mouse Transmembrane Protease, Serine 2 (TMPRSS2) ELISA Kit BYT-ORB780783

Article Name	Mouse Transmembrane Protease, Serine 2 (TMPRSS2) ELISA Kit
Biozol Catalog Number	BYT-ORB780783
Supplier Catalog Number	orb780783
Alternative Catalog Number	BYT-ORB780783-48,BYT-ORB780783-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Mouse
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been precoated with an antibody specific to Transmembrane Protease, Serine 2(TMPRSS2)
Concentration	20 ng/mL
Range	0.32-20 ng/mL
Sensitivity	0.116 ng/mL
UniProt	Q9JIQ8
Samples	Serum, plasma, tissue homogenates and other biological fluids.

**Application Notes** 

Application Notes: standard: 20 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mouse TMPRSS2. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse TMPRSS2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mouse TMPRSS2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Mouse TMPRSS2 in the samples is then determined by comparing the OD of the samples to the standard curve