

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Multiplex Assay Kit for Cluster Of Differentiation 38 (CD38) ,etc. by FLIA (Flow Luminescence Immunoassay) USC-LMA752RA**

Article Name	Multiplex Assay Kit for Cluster Of Differentiation 38 (CD38) ,etc. by FLIA (Flow Luminescence Immunoassay)
Biozol Catalog Number	USC-LMA752RA
Supplier Catalog Number	LMA752Ra
Alternative Catalog Number	USC-LMA752RA-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Rat
Product Description	48T, 96T, 96T*5, 96T*10, 96T*100...
Range	0.2-200ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.067 ng/mL
UniProt	<a href="#">Q64244</a>
Samples	Serum, plasma, tissue homogenates and other biological fluids.
Application Notes	FLIA Kit for Antigen Detection.