

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human CA19-9(Carbohydrate Antigen 19-9) ELISA Kit EBT-ELK1198

| | |
|----------------------------|---|
| Article Name | Human CA19-9(Carbohydrate Antigen 19-9) ELISA Kit |
| Biozol Catalog Number | EBT-ELK1198 |
| Supplier Catalog Number | ELK1198 |
| Alternative Catalog Number | EBT-ELK1198-96, EBT-ELK1198-48, EBT-ELK1198-96X5 |
| Manufacturer | ELK Biotechnology |
| Category | Kits/Assays |
| Species Reactivity | Human |
| Concentration | 50 U/mL |
| Range | 0.79-50 U/mL |
| Sensitivity | 0.29 U/mL |
| Samples | Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids. |

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Tumor immunity, Hepatology, Gastroenterology,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CA19-9. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CA19-9. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CA19-9, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CA19-9 in the samples is then determined by comparing the OD of the samples to the standard curve