

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human NTXI(Cross Linked N-Telopeptide Of Type I Collagen) ELISA Kit EBT-ELK1843

Article Name	Human NTXI(Cross Linked N-Telopeptide Of Type I Collagen) ELISA Kit
Biozol Catalog Number	EBT-ELK1843
Supplier Catalog Number	ELK1843
Alternative Catalog Number	EBT-ELK1843-96, EBT-ELK1843-48, EBT-ELK1843-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	200 ng/mL
Range	3.13-200 ng/mL
Sensitivity	0.94 ng/mL
Samples	serum, plasma, cell lysates, cell culture supernates and other biological fluids

Application Notes

Assay Type: Competitive Inhibition. Assay length: 2h. Research Area: Metabolic pathway, Bone metabolism,. Test principle: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Human NTXI. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human NTXI. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human NTXI in the samples is then determined by comparing the OD of the samples to the standard curve