

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human VIPR1(Vasoactive Intestinal Peptide Receptor 1) ELISA Kit EBT-ELK3082

Article Name	Human VIPR1(Vasoactive Intestinal Peptide Receptor 1) ELISA Kit
Biozol Catalog Number	EBT-ELK3082
Supplier Catalog Number	ELK3082
Alternative Catalog Number	EBT-ELK3082-96, EBT-ELK3082-48, EBT-ELK3082-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	10 ng/mL
Range	0.16-10 ng/mL
Sensitivity	0.054 ng/mL
UniProt	<a href="#">P32241</a>
Samples	Tissue homogenates, cell lysates and other biological fluids.

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Endocrinology,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human VIPR1. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human VIPR1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human VIPR1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human VIPR1 in the samples is then determined by comparing the OD of the samples to the standard curve