

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human CD109(Cluster Of Differentiation 109) ELISA Kit EBT-ELK4321

Article Name	Human CD109(Cluster Of Differentiation 109) ELISA Kit
Biozol Catalog Number	EBT-ELK4321
Supplier Catalog Number	ELK4321
Alternative Catalog Number	EBT-ELK4321-96, EBT-ELK4321-48, EBT-ELK4321-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	10 ng/mL
Range	0.16-10 ng/mL
Sensitivity	0.056 ng/mL
UniProt	Q6YHK3
Samples	Tissue homogenates, cell lysates and other biological fluids

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule, Tumor immunity, Infection immunity, Hematology, Autoimmunity, . Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human CD109. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human CD109. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human CD109, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human CD109 in the samples is then determined by comparing the OD of the samples to the standard curve