

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human HIF1aN(Hypoxia Inducible Factor 1 Alpha Subunit Inhibitor) ELISA Kit EBT-ELK4964

Article Name	Human HIF1aN(Hypoxia Inducible Factor 1 Alpha Subunit Inhibitor) ELISA Kit
Biozol Catalog Number	EBT-ELK4964
Supplier Catalog Number	ELK4964
Alternative Catalog Number	EBT-ELK4964-96, EBT-ELK4964-48, EBT-ELK4964-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	2000 pg/mL
Range	31.25-2000 pg/mL
Sensitivity	14.2 pg/mL
UniProt	<a href="#">Q9NWT6</a>
Samples	Serum, plasma, tissue homogenates, cell lysates and other biological fluids.

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Immune molecule,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human HIF1aN. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human HIF1aN. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human HIF1aN, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human HIF1aN in the samples is then determined by comparing the OD of the samples to the standard curve