

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human SPLUNC2(Short Palate, Lung And Nasal Epithelium Carcinoma Associated Protein 2) ELISA Kit EBT-ELK6899

Article Name	Human SPLUNC2(Short Palate, Lung And Nasal Epithelium Carcinoma Associated Protein 2) ELISA Kit
Biozol Catalog Number	EBT-ELK6899
Supplier Catalog Number	ELK6899
Alternative Catalog Number	EBT-ELK6899-96, EBT-ELK6899-48, EBT-ELK6899-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	10 ng/mL
Range	0.16-10 ng/mL
Sensitivity	0.063 ng/mL
UniProt	Q96DR5
Samples	Serum, plasma, tissue homogenates and other biological fluids.

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: Signal transduction,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human SPLUNC2. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human SPLUNC2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human SPLUNC2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human SPLUNC2 in the samples is then determined by comparing the OD of the samples to the standard curve