

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human LYVE1(Lymphatic Vessel Endothelial Hyaluronan Receptor 1) ELISA Kit EBT-ELK9650

Article Name	Human LYVE1(Lymphatic Vessel Endothelial Hyaluronan Receptor 1) ELISA Kit
Biozol Catalog Number	EBT-ELK9650
Supplier Catalog Number	ELK9650
Alternative Catalog Number	EBT-ELK9650-96, EBT-ELK9650-48, EBT-ELK9650-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	1000 pg/mL
Range	15.63-1000 pg/mL
Sensitivity	6.1 pg/mL
UniProt	Q9Y5Y7
Samples	Serum, plasma, tissue homogenates and other biological fluids.

Application Notes

Assay Type: Sandwich. Assay length: 3.5h. Research Area: CD & Adhesion molecule,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human LYVE1. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human LYVE1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human LYVE1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human LYVE1 in the samples is then determined by comparing the OD of the samples to the standard curve