

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Rat SEMA7A(Semaphorin 7A) ELISA Kit EBT-ELK9667

|                            |   |
|----------------------------|---|
| Article Name               | Rat SEMA7A(Semaphorin 7A) ELISA Kit   |
| Biozol Catalog Number      | EBT-ELK9667   |
| Supplier Catalog Number    | ELK9667   |
| Alternative Catalog Number | EBT-ELK9667-96, EBT-ELK9667-48, EBT-ELK9667-96X5                            |
| Manufacturer               | ELK Biotechnology   |
| Category                   | Kits/Assays   |
| Species Reactivity         | Rat   |
| Concentration              | 20 ng/mL  |
| Range                      | 0.32-20 ng/mL   |
| Sensitivity                | 0.116 ng/mL   |
| UniProt                    | <a href="#">D3ZQP6</a>  |
| Samples                    | serum, plasma, tissue homogenates, cell lysates and other biological fluids |

|                   |   |
|-------------------|---|
| Application Notes | <p>Assay Type: Sandwich. Assay length: 3.5h. Research Area: Signal transduction,CD &amp; Adhesion molecule,Tumor immunity,Infection immunity,Autoimmunity,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Rat SEMA7A. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Rat SEMA7A. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Rat SEMA7A, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Rat SEMA7A in the samples is then determined by comparing the OD of the samples to the standard curve</p> |
|-------------------|---|