

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human Spindlin-2A (SPIN2A) ELISA Kit ASC-KTE60523

|                            |  |
|----------------------------|--|
| Article Name               | Human Spindlin-2A (SPIN2A) ELISA Kit   |
| Biozol Catalog Number      | ASC-KTE60523   |
| Supplier Catalog Number    | KTE60523   |
| Alternative Catalog Number | ASC-KTE60523-48, ASC-KTE60523-96   |
| Manufacturer               | Abbkine Scientific   |
| Category                   | Kits/Assays  |
| Application                | ELISA  |
| Species Reactivity         | Human  |
| Product Description        | This Human Spindlin-2A (SPIN2A) ELISA Kit employs a two-site sandwich ELISA to quantitate SPIN2A.... |
| Range                      | Please inquire   |
| Sensitivity                | Please inquire   |
| Tag                        | SPIN2A   |
| NCBI                       | <a href="#">54466</a>  |
| UniProt                    | <a href="#">Q99865</a>   |
| Samples                    | Cell culture supernatantsSerumPlasmaOther biological fluids  |

Application Notes

This Human Spindlin-2A (SPIN2A) ELISA Kit employs a two-site sandwich ELISA to quantitate SPIN2A in samples. An antibody specific for SPIN2A has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any SPIN2A present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for SPIN2A is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of SPIN2A bound in the initial step. The color development is stopped and the intensity of the color is measured.