

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human Myocardin (MYOCD) ELISA Kit ASC-KTE61409

|                            |  |
|----------------------------|--|
| Article Name               | Human Myocardin (MYOCD) ELISA Kit  |
| Biozol Catalog Number      | ASC-KTE61409   |
| Supplier Catalog Number    | KTE61409   |
| Alternative Catalog Number | ASC-KTE61409-48, ASC-KTE61409-96   |
| Manufacturer               | Abbkine Scientific   |
| Category                   | Kits/Assays  |
| Application                | ELISA  |
| Species Reactivity         | Human  |
| Product Description        | This Human Myocardin (MYOCD) ELISA Kit employs a two-site sandwich ELISA to quantitate MYOCD.... |
| Range                      | Please inquire   |
| Sensitivity                | Please inquire   |
| Tag                        | MYOCD  |
| NCBI                       | <a href="#">93649</a>  |
| UniProt                    | <a href="#">Q8IZQ8</a>   |
| Samples                    | Cell culture supernatantsSerumPlasmaOther biological fluids                                      |

Application Notes

This Human Myocardin (MYOCD) ELISA Kit employs a two-site sandwich ELISA to quantitate MYOCD in samples. An antibody specific for MYOCD has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MYOCD present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MYOCD is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MYOCD bound in the initial step. The color development is stopped and the intensity of the color is measured.