BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Matrix metalloproteinase 8/Neutrophil collagenase (MMP8) ELISA Kit
ASC-KTE61585

Article Name	Human Matrix metalloproteinase 8/Neutrophil collagenase (MMP8) ELISA Kit
Biozol Catalog Number	ASC-KTE61585
Supplier Catalog Number	KTE61585
Alternative Catalog Number	ASC-KTE61585-48, ASC-KTE61585-96
Manufacturer	Abbkine Scientific
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	Human Matrix metalloproteinase 8/Neutrophil collagenase (MMP8) ELISA Kit employs a two-site sandwich ELISA to quantitate MMP8 in samples....
Range	15 µg/L-240 µg/L
Sensitivity	1.0 µg/L
Tag	MMP8
Samples	Cell culture supernatantsSerumPlasmaOther biological fluids

Application Notes	Human Matrix metalloproteinase 8/Neutrophil collagenase (MMP8) ELISA Kit employs a two-site sandwich ELISA to quantitate MMP8 in samples. An antibody specific for MMP8 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MMP8 present is bound by the immobilized antibody. After removing any unbound substances, HRP-Conjugate Human MMP8 detection antibody is added to the wells. Following a wash to remove any unbound HRP reagent, a Chromogen solution is added to the wells and color develops in proportion to the amount of MMP8 bound in the initial step. The color development is stopped and the intensity of the color is measured.