BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Lymphatic vessel endothelial hyaluronic acid receptor 1 (LYVE1) ELISA Kit
ASC-KTE61763

Article Name	Human Lymphatic vessel endothelial hyaluronic acid receptor 1 (LYVE1) ELISA Kit
Biozol Catalog Number	ASC-KTE61763
Supplier Catalog Number	KTE61763
Alternative Catalog Number	ASC-KTE61763-48, ASC-KTE61763-96
Manufacturer	Abbkine Scientific
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	This Human Lymphatic vessel endothelial hyaluronic acid receptor 1 (LYVE1) ELISA Kit employs a two-site sandwich ELISA to quantitate LYVE1....
Range	Please inquire
Sensitivity	Please inquire
Tag	LYVE1
NCBI	10894
UniProt	Q9Y5Y7
Samples	Cell culture supernatantsSerumPlasmaOther biological fluids

Application Notes	This Human Lymphatic vessel endothelial hyaluronic acid receptor 1 (LYVE1) ELISA Kit employs a two-site sandwich ELISA to quantitate LYVE1 in samples. An antibody specific for LYVE1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyLYVE1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for LYVE1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of LYVE1 bound in the initial step. The color development is stopped and the intensity of the color is measured.