

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Anti-Rotavirus IgG ELISA Kit ASC-KTE63064

| | |
|----------------------------|---|
| Article Name | Human Anti-Rotavirus IgG ELISA Kit |
| Biozol Catalog Number | ASC-KTE63064 |
| Supplier Catalog Number | KTE63064 |
| Alternative Catalog Number | ASC-KTE63064-48, ASC-KTE63064-96 |
| Manufacturer | Abbkine Scientific |
| Category | Kits/Assays |
| Application | ELISA |
| Species Reactivity | Human |
| Product Description | Human Anti-Rotavirus IgG ELISA Kit quantitate Human Anti-Rotavirus IgG in samples.... |
| Range | 1.5 ng/L-24 ng/L |
| Sensitivity | 0.1ng/L |
| Tag | Anti-Rotavirus IgG |
| Samples | Cell culture supernatantsSerumPlasmaOther biological fluids |

Application Notes

Human Anti-Rotavirus IgG ELISA Kit quantitate Human Anti-Rotavirus IgG in samples. An antigen has been pre-coated onto a microplate. Standards or test samples are added to the wells, incubated and then washed. HRP-conjugated anti-human IgG antibody is then added and incubated. The plate is washed once more and the chromogen solution is then added which HRP catalyzes, generating a blue coloration after incubation. A stop solution is added which generates conversion to yellow color read at 450 nm which is proportional to the amount of analyte bound.