BIOZOL

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Mouse intercellular adhesion molecule 1 (ICAM-1/CD54) Microsample Fast ELISA Kit
BYT-ORB1806693

Article Name	Mouse intercellular adhesion molecule 1 (ICAM-1/CD54) Microsample Fast ELISA Kit
Biozol Catalog Number	BYT-ORB1806693
Supplier Catalog Number	orb1806693
Alternative Catalog Number	BYT-ORB1806693-48,BYT-ORB1806693-96
Manufacturer	Biorbyt
Category	Kits/Assays
Species Reactivity	Mouse
Product Description	Mouse intercellular adhesion molecule 1 (ICAM-1/CD54) Microsample Fast ELISA Kit...
Range	1.37-1000 ng/mL
Sensitivity	0.82 ng/mL
UniProt	P13597
Form	Ready to use

Samples	serum, plasma and other biological fluids, Sample volume:100μL,Sample volume:25μL
Target	ICAM-1/CD54
Application Notes	Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Mouse ICAM-1/CD54. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Mouse ICAM-1/CD54 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Mouse ICAM-1/CD54, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm 2 nm. The OD value is proportional to the concentration of Mouse ICAM-1/CD54. You can calculate the concentration of Mouse ICAM-1/CD54 in the samples by comparing the OD of the samples to the standard curve