

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Flotillin 2 (FLOT2) ELISA Kit BYT-ORB1949793

Article Name	Human Flotillin 2 (FLOT2) ELISA Kit
Biozol Catalog Number	BYT-ORB1949793
Supplier Catalog Number	orb1949793
Alternative Catalog Number	BYT-ORB1949793-48, BYT-ORB1949793-96
Manufacturer	Biorbyt
Category	Kits/Assays
Species Reactivity	Human
Product Description	Human Flotillin 2 (FLOT2) ELISA Kit...
Range	0.63-40ng/mL
Sensitivity	0.38 ng/mL
UniProt	Q14254
Samples	Tissue homogenate
Target	FLOT2

Application Notes

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human FLOT2. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human FLOT2 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human FLOT2, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Human FLOT2. You can calculate the concentration of Human FLOT2 in the samples by comparing the OD of the samples to the standard curve