

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Rabbit Selectin, Endothelium (SELE) ELISA Kit BYT-ORB776346

Article Name	Rabbit Selectin, Endothelium (SELE) ELISA Kit
Biozol Catalog Number	BYT-ORB776346
Supplier Catalog Number	orb776346
Alternative Catalog Number	BYT-ORB776346-48, BYT-ORB776346-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Rabbit
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Selectin, Endothelium(SELE)....
Concentration	400 ng/mL
Range	6.25-400 ng/mL
Sensitivity	2.24 ng/mL
UniProt	P27113
Samples	Serum, plasma and other biological fluids.

Application Notes

Application Notes: standard: 400 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Rabbit SELE. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Rabbit SELE. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Rabbit SELE, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Rabbit SELE in the samples is then determined by comparing the OD of the samples to the standard curve