

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Inducible T-Cell Co Stimulator (ICOS) ELISA Kit BYT-ORB776998

Article Name	Human Inducible T-Cell Co Stimulator (ICOS) ELISA Kit
Biozol Catalog Number	BYT-ORB776998
Supplier Catalog Number	orb776998
Alternative Catalog Number	BYT-ORB776998-48, BYT-ORB776998-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Inducible T-Cell Co Stimulator(ICOS)....
Concentration	1000 pg/mL
Range	15.63-1000 pg/mL
Sensitivity	5.6 pg/mL
UniProt	Q9Y6W8
Samples	serum, plasma, urine, cell lysates and other biological fluids

Application Notes

Application Notes: standard: 1000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human ICOS. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human ICOS. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human ICOS, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human ICOS in the samples is then determined by comparing the OD of the samples to the standard curve