

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Rat A Disintegrin and Metalloprotease 10 (ADAM10) ELISA Kit BYT-ORB777008**

|                            |  |
|----------------------------|--|
| Article Name               | Rat A Disintegrin and Metalloprotease 10 (ADAM10) ELISA Kit  |
| Biozol Catalog Number      | BYT-ORB777008  |
| Supplier Catalog Number    | orb777008  |
| Alternative Catalog Number | BYT-ORB777008-48, BYT-ORB777008-96   |
| Manufacturer               | Biorbyt  |
| Category                   | Kits/Assays  |
| Application                | ELISA  |
| Species Reactivity         | Rat  |
| Product Description        | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to A Disintegrin And Metalloprotease 10(ADAM10).... |
| Concentration              | 5000 pg/mL   |
| Range                      | 78.13-5000 pg/mL   |
| Sensitivity                | 27 pg/mL   |
| UniProt                    | <a href="#">Q10743</a>   |
| Samples                    | Serum, plasma, tissue homogenates and other biological fluids  |

Application Notes

Application Notes: standard: 5000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Rat ADAM10. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Rat ADAM10. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Rat ADAM10, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Rat ADAM10 in the samples is then determined by comparing the OD of the samples to the standard curve