

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Placenta Specific Protein 9 (PLAC9) ELISA Kit BYT-ORB780741

Article Name	Human Placenta Specific Protein 9 (PLAC9) ELISA Kit
Biozol Catalog Number	BYT-ORB780741
Supplier Catalog Number	orb780741
Alternative Catalog Number	BYT-ORB780741-48, BYT-ORB780741-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Placenta Specific Protein 9(PLAC9)....
Concentration	50 ng/mL
Range	0.79-50 ng/mL
Sensitivity	0.29 ng/mL
UniProt	Q5JTB6
Samples	Tissue homogenates, cell lysates, cell culture supernates and other biological fluids.

Application Notes

Application Notes: standard: 50 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human PLAC9. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human PLAC9. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human PLAC9, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human PLAC9 in the samples is then determined by comparing the OD of the samples to the standard curve