

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human Angiotensin 1-9 (Ang1-9) ELISA Kit BYT-ORB781452

Article Name	Human Angiotensin 1-9 (Ang1-9) ELISA Kit
Biozol Catalog Number	BYT-ORB781452
Supplier Catalog Number	orb781452
Alternative Catalog Number	BYT-ORB781452-48, BYT-ORB781452-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Ang1-9(Angiotensin 1-9)....
Concentration	1000 pg/mL
Range	15.63-1000 pg/mL
Sensitivity	5.31 pg/mL
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

Application Notes

Application Notes: standard: 1000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human Ang1-9. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human Ang1-9. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human Ang1-9, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human Ang1-9 in the samples is then determined by comparing the OD of the samples to the standard curve