

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Protein Induced by Vitamin K Absence or Antagonist-II (PIVKA-II) ELISA Kit BYT-ORB782364

Article Name	Human Protein Induced by Vitamin K Absence or Antagonist-II (PIVKA-II) ELISA Kit
Biozol Catalog Number	BYT-ORB782364
Supplier Catalog Number	orb782364
Alternative Catalog Number	BYT-ORB782364-48, BYT-ORB782364-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to PIVKA-II(Protein Induced Vitamin K Absence or Antagonist-II)....
Concentration	20 ng/mL
Range	0.32-20 ng/mL
Sensitivity	0.129 ng/mL
Samples	serum, plasma and other biological fluids

Application Notes

Application Notes: standard: 20 ng/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human PIVKA-II. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human PIVKA-II. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human PIVKA-II, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human PIVKA-II in the samples is then determined by comparing the OD of the samples to the standard curve