

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human Procollagen I (PCI) ELISA Kit BYT-ORB782499

Article Name	Human Procollagen I (PCI) ELISA Kit
Biozol Catalog Number	BYT-ORB782499
Supplier Catalog Number	orb782499
Alternative Catalog Number	BYT-ORB782499-48, BYT-ORB782499-96
Manufacturer	Biorbyt
Category	Kits/Assays
Application	ELISA
Species Reactivity	Human
Product Description	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Procollagen I(PCI)....
Concentration	5000 pg/mL
Range	78.13-5000 pg/mL
Sensitivity	26.7 pg/mL
Samples	Serum, plasma and other biological fluids.

Application Notes

Application Notes: standard: 5000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human PCI. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human PCI. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human PCI, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human PCI in the samples is then determined by comparing the OD of the samples to the standard curve