

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human Transmembrane Protein 48 (TMEM48) ELISA Kit BYT-ORB782607

|                            |  |
|----------------------------|--|
| Article Name               | Human Transmembrane Protein 48 (TMEM48) ELISA Kit  |
| Biozol Catalog Number      | BYT-ORB782607  |
| Supplier Catalog Number    | orb782607  |
| Alternative Catalog Number | BYT-ORB782607-48, BYT-ORB782607-96   |
| Manufacturer               | Biorbyt  |
| Category                   | Kits/Assays  |
| Application                | ELISA  |
| Species Reactivity         | Human  |
| Product Description        | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to TMEM48(Transmembrane protein 48).... |
| Concentration              | 2000 pg/mL   |
| Range                      | 31.25-2000 pg/mL   |
| Sensitivity                | 11.9 pg/mL   |
| UniProt                    | <a href="#">Q9BTX1</a>   |
| Samples                    | serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids   |

Application Notes

Application Notes: standard: 2000 pg/mL. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human TMEM48. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human TMEM48. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human TMEM48, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human TMEM48 in the samples is then determined by comparing the OD of the samples to the standard curve