

Diagnostica Vertrieb GmbH, Oehleckerring 11-13

22419 Hamburg, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Trichomonas vaginalis PCR kit (Form T) - CE IVD ATB-84-06

Article Name	Trichomonas vaginalis PCR kit (Form T) - CE IVD
Biozol Catalog Number	ATB-84-06
Supplier Catalog Number	84-06
Alternative Catalog Number	ATB-84-06
Manufacturer	Astra Biotech GmbH
Category	Kits/Assays
Application	Real time PCR based diagnostic kit
Species Reactivity	Human
Product Description	Real time PCR based diagnostic kit for qualitative determination of Trichomonas vaginalis DNA (in tubes)
Sensitivity	Analytical sensitivity: 480 copies of Trichomonas vaginalis genomic DNA / 1 mL
Samples	Urethral swabs, endocervical and vaginal smears, urine
Target	Trichomonas vaginalis
Application Notes	Format T: PCR mix is aliquoted into standard individual tubes Validated with Rotor-Gene Q /Rotor-Gene 3000/6000 (Qiagen/Corbett Research) All required components included and ready to use (PCR mix+ DNA Polymerase +Internal control, Positive control, Nega