



Clinical Study for Rapid Response™ COVID-19 Antigen Rapid Test Cassette (Oral Fluid)

The Rapid Response™ COVID-19 Antigen Rapid Test Cassette has been evaluated with saliva specimens. The study was completed at multiple clinical sites in the United States. A highly sensitive US FDA EUA approved PCR test was used as the comparator method for this evaluation.

Reagent: Rapid Response™ COVID-19 Antigen Rapid Test Cassette
 Product code: COV-2C25B
 Lot numbers: COVG2007001-T, Exp:2022-07
 COVG2007002-T, Exp:2022-07
 COVG2007003-T, Exp:2022-07
 PCR: Thermo Fisher TaqPath COVID-19 Combo kit

Method:

Study 1: Each individual was sampled with a Nasopharyngeal swab and a Saliva specimen. The NP swab was sent for the PCR test. The saliva sample collection and testing procedure was followed from the Product Insert of COV-2C25B. Specimens were considered positive when the result of PCR was positive.

Clinical Study summary of Rapid Response™ COVID-19 Antigen Test Cassette for Study 1:

COVG2007001-T, Exp:2022-07
 COVG2007002-T, Exp:2022-07
 COVG2007003-T, Exp:2022-07

Method		PCR (NP swab)		Total Results
		Positive	Negative	
Antigen Rapid Test Cassette (Saliva)	Results			
	Positive	38	0	38
	Negative	2	324	326
Total Results		40	324	364

Relative Sensitivity: 95.0% (95%CI*:83.1%-99.1%)*

Relative Specificity: 100% (95%CI*:98.53%-100.0%)*

Relative accuracy: 99.5% (95%CI*:97.8%-99.9%)*

Study 2: The study was extended to evaluate the performance on frozen specimen. The Nasopharyngeal swab was collected for the PCR and saliva specimen from each individual was collected to be stored at -20°C. The saliva specimens of the PCR positive individuals was tested

with the Rapid Response™ COVID-19 Antigen Rapid Test Cassette after bringing it to room temperature.

Clinical Study summary of Rapid Response™ COVID-19 Antigen Test Cassette for Study 2:
COVG2007001-T, Exp:2022-07

Method		PCR (NP swab)		Total Results
Antigen Rapid Test Cassette (Frozen Saliva)	Results	Positive	Negative	
	Positive	88	0	88
	Negative	11	0	11
Total Results		99	0	99

Relative Sensitivity: 88.9% (95%CI*:80.6%-94.1%)*

Combined Clinical Results of Study 1 and Study 2:

Method		PCR (NP swab)		Total Results
Antigen Rapid Test Cassette (Saliva)	Results	Positive	Negative	
	Positive	126	0	126
	Negative	13	324	337
Total Results		139	324	463

Relative Sensitivity: 90.6% (95%CI*:84.5%-94.9%)*

Relative Specificity: 100% (95%CI*:98.53%-100.0%)*

Relative accuracy: 99.5% (95%CI*:97.8%-99.9%)*

Clinical Data from Study 1:

	Specimen	PCR Comparator	Rapid Response™ COVID-19 Antigen Rapid Test Cassette(Oral Fluid)	Gender	Age
1	Oral fluid	Negative	Negative	Male	45
2	Oral fluid	Negative	Negative	Male	57
3	Oral fluid	Negative	Negative	Female	69
4	Oral fluid	Negative	Negative	Female	64
5	Oral fluid	Negative	Negative	Female	42
6	Oral fluid	Negative	Negative	Male	52
7	Oral fluid	Positive	Positive	Female	50
8	Oral fluid	Negative	Negative	Male	56
9	Oral fluid	Negative	Negative	Female	64
10	Oral fluid	Positive	Positive	Male	42
11	Oral fluid	Negative	Negative	Male	35
12	Oral fluid	Negative	Negative	Male	65



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

13	Oral fluid	Negative	Negative	Male	67
14	Oral fluid	Negative	Negative	Male	50
15	Oral fluid	Negative	Negative	Male	36
16	Oral fluid	Negative	Negative	Male	43
17	Oral fluid	Negative	Negative	Male	61
18	Oral fluid	Negative	Negative	Female	56
19	Oral fluid	Negative	Negative	Female	55
20	Oral fluid	Negative	Negative	Female	37
21	Oral fluid	Negative	Negative	Male	63
22	Oral fluid	Negative	Negative	Male	49
23	Oral fluid	Negative	Negative	Male	56
24	Oral fluid	Positive	Positive	Female	20
25	Oral fluid	Negative	Negative	Female	61
26	Oral fluid	Negative	Negative	Male	54
27	Oral fluid	Negative	Negative	Female	56
28	Oral fluid	Positive	Positive	Male	38
29	Oral fluid	Negative	Negative	Female	50
30	Oral fluid	Negative	Negative	Male	52
31	Oral fluid	Positive	Positive	Female	39
32	Oral fluid	Negative	Negative	Male	71
33	Oral fluid	Negative	Negative	Female	72
34	Oral fluid	Negative	Negative	Male	58
35	Oral fluid	Negative	Negative	Male	71
36	Oral fluid	Negative	Negative	Male	68
37	Oral fluid	Positive	Positive	Female	34
38	Oral fluid	Negative	Negative	Male	54
39	Oral fluid	Negative	Negative	Female	72
40	Oral fluid	Negative	Negative	Female	45
41	Oral fluid	Negative	Negative	Male	63
42	Oral fluid	Negative	Negative	Male	65
43	Oral fluid	Negative	Negative	Male	64
44	Oral fluid	Negative	Negative	Female	68
45	Oral fluid	Positive	Positive	Female	51
46	Oral fluid	Negative	Negative	Male	61
47	Oral fluid	Negative	Negative	Male	40
48	Oral fluid	Negative	Negative	Male	49
49	Oral fluid	Negative	Negative	Male	64
50	Oral fluid	Negative	Negative	Female	59
51	Oral fluid	Negative	Negative	Female	49
52	Oral fluid	Negative	Negative	Female	41
53	Oral fluid	Negative	Negative	Female	36
54	Oral fluid	Negative	Negative	Female	63
55	Oral fluid	Negative	Negative	Male	67
56	Oral fluid	Negative	Negative	Female	71
57	Oral fluid	Negative	Negative	Male	38
58	Oral fluid	Positive	Positive	Female	67



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

59	Oral fluid	Negative	Negative	Female	69
60	Oral fluid	Negative	Negative	Female	50
61	Oral fluid	Negative	Negative	Male	64
62	Oral fluid	Negative	Negative	Female	37
63	Oral fluid	Negative	Negative	Female	69
64	Oral fluid	Negative	Negative	Female	56
65	Oral fluid	Negative	Negative	Male	43
66	Oral fluid	Negative	Negative	Female	71
67	Oral fluid	Negative	Negative	Female	49
68	Oral fluid	Negative	Negative	Male	37
69	Oral fluid	Negative	Negative	Male	45
70	Oral fluid	Negative	Negative	Male	44
71	Oral fluid	Negative	Negative	Female	56
72	Oral fluid	Negative	Negative	Female	43
73	Oral fluid	Negative	Negative	Male	42
74	Oral fluid	Negative	Negative	Male	64
75	Oral fluid	Negative	Negative	Female	59
76	Oral fluid	Negative	Positive	Male	56
77	Oral fluid	Positive	Positive	Male	70
78	Oral fluid	Negative	Negative	Male	57
79	Oral fluid	Negative	Negative	Male	39
80	Oral fluid	Negative	Negative	Male	44
81	Oral fluid	Negative	Negative	Female	35
82	Oral fluid	Negative	Negative	Male	68
83	Oral fluid	Negative	Negative	Male	50
84	Oral fluid	Positive	Positive	Female	27
85	Oral fluid	Negative	Negative	Female	63
86	Oral fluid	Negative	Negative	Male	37
87	Oral fluid	Negative	Negative	Male	68
88	Oral fluid	Negative	Negative	Female	59
89	Oral fluid	Negative	Negative	Female	65
90	Oral fluid	Negative	Negative	Male	72
91	Oral fluid	Negative	Negative	Male	53
92	Oral fluid	Negative	Negative	Male	45
93	Oral fluid	Negative	Negative	Male	45
94	Oral fluid	Negative	Negative	Female	49
95	Oral fluid	Positive	Positive	Male	54
96	Oral fluid	Negative	Negative	Male	46
97	Oral fluid	Negative	Negative	Male	35
98	Oral fluid	Negative	Negative	Male	45
99	Oral fluid	Negative	Negative	Female	66
100	Oral fluid	Negative	Negative	Male	40
101	Oral fluid	Negative	Negative	Female	68
102	Oral fluid	Negative	Negative	Female	42
103	Oral fluid	Negative	Negative	Male	53
104	Oral fluid	Negative	Negative	Female	35



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

105	Oral fluid	Negative	Negative	Female	67
106	Oral fluid	Negative	Negative	Female	70
107	Oral fluid	Negative	Negative	Male	51
108	Oral fluid	Negative	Negative	Female	72
109	Oral fluid	Negative	Negative	Female	58
110	Oral fluid	Negative	Negative	Male	62
111	Oral fluid	Negative	Negative	Female	53
112	Oral fluid	Negative	Negative	Female	38
113	Oral fluid	Negative	Negative	Female	41
114	Oral fluid	Negative	Negative	Female	46
115	Oral fluid	Negative	Negative	Female	66
116	Oral fluid	Negative	Negative	Female	47
117	Oral fluid	Negative	Negative	Female	67
118	Oral fluid	Negative	Negative	Male	68
119	Oral fluid	Negative	Negative	Male	43
120	Oral fluid	Negative	Negative	Female	50
121	Oral fluid	Negative	Negative	Male	38
122	Oral fluid	Negative	Negative	Female	37
123	Oral fluid	Negative	Negative	Male	38
124	Oral fluid	Negative	Negative	Female	66
125	Oral fluid	Negative	Negative	Male	35
126	Oral fluid	Negative	Negative	Male	42
127	Oral fluid	Negative	Negative	Male	54
128	Oral fluid	Positive	Positive	Female	22
129	Oral fluid	Negative	Negative	Male	70
130	Oral fluid	Negative	Negative	Male	57
131	Oral fluid	Negative	Negative	Female	37
132	Oral fluid	Negative	Negative	Female	36
133	Oral fluid	Negative	Negative	Male	54
134	Oral fluid	Negative	Negative	Male	56
135	Oral fluid	Negative	Negative	Male	56
136	Oral fluid	Negative	Negative	Female	71
137	Oral fluid	Negative	Negative	Female	52
138	Oral fluid	Negative	Negative	Male	69
139	Oral fluid	Negative	Negative	Male	41
140	Oral fluid	Negative	Negative	Male	36
141	Oral fluid	Negative	Negative	Female	70
142	Oral fluid	Negative	Negative	Male	69
143	Oral fluid	Negative	Negative	Male	49
144	Oral fluid	Negative	Negative	Female	68
145	Oral fluid	Negative	Negative	Male	35
146	Oral fluid	Negative	Negative	Female	69
147	Oral fluid	Negative	Negative	Female	41
148	Oral fluid	Negative	Negative	Female	49
149	Oral fluid	Negative	Negative	Male	37
150	Oral fluid	Negative	Negative	Female	61



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

151	Oral fluid	Negative	Negative	Male	60
152	Oral fluid	Negative	Negative	Female	56
153	Oral fluid	Negative	Negative	Male	49
154	Oral fluid	Negative	Negative	Female	36
155	Oral fluid	Positive	Positive	Female	36
156	Oral fluid	Negative	Negative	Female	64
157	Oral fluid	Negative	Negative	Male	58
158	Oral fluid	Negative	Negative	Male	67
159	Oral fluid	Negative	Negative	Male	50
160	Oral fluid	Negative	Negative	Male	54
161	Oral fluid	Negative	Negative	Female	35
162	Oral fluid	Negative	Negative	Male	47
163	Oral fluid	Negative	Negative	Male	51
164	Oral fluid	Negative	Negative	Female	58
165	Oral fluid	Negative	Negative	Male	41
166	Oral fluid	Negative	Negative	Male	38
167	Oral fluid	Negative	Negative	Male	41
168	Oral fluid	Negative	Negative	Male	39
169	Oral fluid	Negative	Negative	Male	44
170	Oral fluid	Negative	Negative	Male	49
171	Oral fluid	Negative	Negative	Male	42
172	Oral fluid	Negative	Negative	Female	44
173	Oral fluid	Negative	Negative	Male	35
174	Oral fluid	Negative	Negative	Female	41
175	Oral fluid	Negative	Negative	Female	44
176	Oral fluid	Negative	Negative	Male	56
177	Oral fluid	Negative	Negative	Male	55
178	Oral fluid	Negative	Negative	Female	39
179	Oral fluid	Negative	Negative	Male	49
180	Oral fluid	Negative	Negative	Female	39
181	Oral fluid	Negative	Negative	Female	44
182	Oral fluid	Negative	Negative	Male	59
183	Oral fluid	Negative	Negative	Female	56
184	Oral fluid	Negative	Negative	Male	44
185	Oral fluid	Negative	Negative	Male	50
186	Oral fluid	Negative	Negative	Male	36
187	Oral fluid	Negative	Negative	Male	47
188	Oral fluid	Negative	Negative	Male	38
189	Oral fluid	Negative	Negative	Female	52
190	Oral fluid	Negative	Negative	Male	36
191	Oral fluid	Negative	Negative	Female	43
192	Oral fluid	Negative	Negative	Female	59
193	Oral fluid	Negative	Negative	Male	68
194	Oral fluid	Negative	Negative	Male	60
195	Oral fluid	Positive	Positive	Female	47
196	Oral fluid	Negative	Negative	Male	57



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

197	Oral fluid	Negative	Negative	Male	45
198	Oral fluid	Negative	Negative	Female	70
199	Oral fluid	Negative	Negative	Female	69
200	Oral fluid	Positive	Positive	Male	62
201	Oral fluid	Negative	Negative	Male	72
202	Oral fluid	Negative	Negative	Female	72
203	Oral fluid	Negative	Negative	Female	41
204	Oral fluid	Negative	Negative	Female	61
205	Oral fluid	Positive	Positive	Female	52
206	Oral fluid	Negative	Negative	Male	54
207	Oral fluid	Negative	Negative	Female	53
208	Oral fluid	Negative	Negative	Female	71
209	Oral fluid	Negative	Negative	Male	39
210	Oral fluid	Negative	Negative	Female	65
211	Oral fluid	Negative	Negative	Male	40
212	Oral fluid	Positive	Positive	Male	41
213	Oral fluid	Negative	Negative	Female	46
214	Oral fluid	Negative	Negative	Female	70
215	Oral fluid	Negative	Negative	Male	52
216	Oral fluid	Negative	Negative	Male	57
217	Oral fluid	Negative	Negative	Male	54
218	Oral fluid	Negative	Negative	Female	41
219	Oral fluid	Positive	Positive	Male	49
220	Oral fluid	Negative	Negative	Female	56
221	Oral fluid	Negative	Negative	Male	57
222	Oral fluid	Positive	Positive	Female	50
223	Oral fluid	Negative	Negative	Male	71
224	Oral fluid	Negative	Negative	Female	45
225	Oral fluid	Negative	Negative	Female	57
226	Oral fluid	Negative	Negative	Female	35
227	Oral fluid	Negative	Negative	Female	58
228	Oral fluid	Negative	Negative	Female	67
229	Oral fluid	Negative	Negative	Male	35
230	Oral fluid	Negative	Positive	Male	42
231	Oral fluid	Negative	Negative	Male	36
232	Oral fluid	Negative	Negative	Female	60
233	Oral fluid	Negative	Negative	Female	68
234	Oral fluid	Negative	Negative	Male	66
235	Oral fluid	Negative	Negative	Male	49
236	Oral fluid	Negative	Negative	Female	56
237	Oral fluid	Negative	Negative	Female	36
238	Oral fluid	Negative	Negative	Female	35
239	Oral fluid	Negative	Negative	Female	72
240	Oral fluid	Negative	Negative	Female	39
241	Oral fluid	Negative	Negative	Female	50
242	Oral fluid	Negative	Negative	Male	72



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

243	Oral fluid	Negative	Negative	Male	64
244	Oral fluid	Negative	Negative	Female	43
245	Oral fluid	Negative	Negative	Male	60
246	Oral fluid	Negative	Negative	Male	68
247	Oral fluid	Negative	Negative	Male	67
248	Oral fluid	Negative	Negative	Female	63
249	Oral fluid	Negative	Negative	Female	53
250	Oral fluid	Negative	Negative	Female	54
251	Oral fluid	Negative	Negative	Female	53
252	Oral fluid	Negative	Negative	Female	39
253	Oral fluid	Negative	Negative	Male	59
254	Oral fluid	Negative	Negative	Male	48
255	Oral fluid	Negative	Negative	Female	44
256	Oral fluid	Negative	Negative	Male	70
257	Oral fluid	Negative	Negative	Female	52
258	Oral fluid	Negative	Negative	Female	54
259	Oral fluid	Negative	Negative	Male	67
260	Oral fluid	Negative	Negative	Male	36
261	Oral fluid	Positive	Negative	Male	46
262	Oral fluid	Negative	Negative	Male	35
263	Oral fluid	Negative	Negative	Female	36
264	Oral fluid	Positive	Positive	Male	63
265	Oral fluid	Positive	Positive	Male	16
266	Oral fluid	Negative	Negative	Male	65
267	Oral fluid	Negative	Negative	Male	52
268	Oral fluid	Negative	Negative	Male	44
269	Oral fluid	Negative	Negative	Female	48
270	Oral fluid	Negative	Negative	Female	64
271	Oral fluid	Negative	Negative	Female	52
272	Oral fluid	Positive	Positive	Female	69
273	Oral fluid	Negative	Negative	Female	68
274	Oral fluid	Negative	Negative	Male	48
275	Oral fluid	Positive	Positive	Female	55
276	Oral fluid	Negative	Negative	Female	58
277	Oral fluid	Negative	Negative	Female	40
278	Oral fluid	Negative	Negative	Female	63
279	Oral fluid	Negative	Negative	Male	71
280	Oral fluid	Negative	Negative	Male	61
281	Oral fluid	Negative	Negative	Male	54
282	Oral fluid	Negative	Negative	Male	70
283	Oral fluid	Negative	Negative	Male	71
284	Oral fluid	Negative	Negative	Male	47
285	Oral fluid	Negative	Negative	Male	66
286	Oral fluid	Negative	Negative	Female	36
287	Oral fluid	Negative	Negative	Male	60
288	Oral fluid	Negative	Negative	Male	47



570 Hood Road, Unit 23,
Markham, ON
T: 905-944-9565
F: 905-944-0406

Clinical Study
Rapid Response™ COVID-19 Antigen Rapid Test Cassette
(Oral Fluid)
COV-2C25B

289	Oral fluid	Negative	Negative	Male	53
290	Oral fluid	Negative	Negative	Female	56
291	Oral fluid	Negative	Negative	Female	67
292	Oral fluid	Negative	Negative	Male	71
293	Oral fluid	Negative	Negative	Female	61
294	Oral fluid	Negative	Negative	Male	71
295	Oral fluid	Negative	Negative	Female	61
296	Oral fluid	Negative	Negative	Female	49
297	Oral fluid	Negative	Negative	Female	64
298	Oral fluid	Negative	Negative	Male	55
299	Oral fluid	Negative	Negative	Female	50
300	Oral fluid	Negative	Negative	Female	71
301	Oral fluid	Negative	Negative	Male	70
302	Oral fluid	Negative	Negative	Female	53
303	Oral fluid	Negative	Negative	Male	61
304	Oral fluid	Negative	Negative	Female	42
305	Oral fluid	Negative	Negative	Male	58
306	Oral fluid	Negative	Negative	Female	46
307	Oral fluid	Negative	Negative	Male	44
308	Oral fluid	Positive	Positive	Female	34
309	Oral fluid	Negative	Negative	Male	41
310	Oral fluid	Negative	Negative	Male	54
311	Oral fluid	Negative	Negative	Female	51
312	Oral fluid	Negative	Negative	Male	53
313	Oral fluid	Negative	Negative	Male	54
314	Oral fluid	Negative	Negative	Male	36
315	Oral fluid	Negative	Negative	Female	61
316	Oral fluid	Negative	Negative	Female	44
317	Oral fluid	Negative	Negative	Male	52
318	Oral fluid	Negative	Negative	Female	53
319	Oral fluid	Negative	Negative	Female	35
320	Oral fluid	Negative	Negative	Female	44
321	Oral fluid	Negative	Negative	Female	56
322	Oral fluid	Negative	Negative	Female	57
323	Oral fluid	Negative	Negative	Female	52
324	Oral fluid	Negative	Negative	Male	36
325	Oral fluid	Negative	Negative	Female	69
326	Oral fluid	Negative	Negative	Male	37
327	Oral fluid	Negative	Negative	Male	54
328	Oral fluid	Negative	Negative	Female	59
329	Oral fluid	Negative	Negative	Male	70
330	Oral fluid	Positive	Positive	Female	28
331	Oral fluid	Positive	Positive	Male	36
332	Oral fluid	Negative	Negative	Female	72
333	Oral fluid	Negative	Negative	Male	63
334	Oral fluid	Negative	Negative	Male	71

335	Oral fluid	Positive	Positive	Female	70
336	Oral fluid	Negative	Negative	Female	61
337	Oral fluid	Negative	Negative	Female	35
338	Oral fluid	Negative	Negative	Male	63
339	Oral fluid	Negative	Negative	Male	59
340	Oral fluid	Negative	Negative	Female	57
341	Oral fluid	Negative	Negative	Female	66
342	Oral fluid	Negative	Negative	Female	51
343	Oral fluid	Negative	Negative	Female	54
344	Oral fluid	Negative	Negative	Male	64
345	Oral fluid	Negative	Negative	Female	59
346	Oral fluid	Negative	Negative	Female	45
347	Oral fluid	Negative	Negative	Male	46
348	Oral fluid	Negative	Negative	Female	57
349	Oral fluid	Negative	Negative	Male	50
350	Oral fluid	Positive	Positive	Female	62
351	Oral fluid	Negative	Negative	Female	52
352	Oral fluid	Negative	Negative	Male	69
353	Oral fluid	Negative	Negative	Male	66
354	Oral fluid	Positive	Negative	Female	69
355	Oral fluid	Positive	Positive	Female	13
356	Oral fluid	Positive	Positive	Male	58
357	Oral fluid	Positive	Positive	Female	35
358	Oral fluid	Positive	Positive	Female	41
359	Oral fluid	Positive	Positive	Male	38
360	Oral fluid	Positive	Positive	Female	42
361	Oral fluid	Positive	Positive	Female	53
362	Oral fluid	Positive	Positive	Male	62
363	Oral fluid	Positive	Positive	Male	29
364	Oral fluid	Positive	Positive	Male	33

Clinical Data from Study 2:

Age	Gender	Symptom Date	Collection Date	Specimen Type	PCR Result	Rapid Test Results
52	M	2020-11-15	2020-11-16	Saliva	Positive	Negative
26	F	2020-11-18	2020-11-19	Saliva	Positive	Negative
40	M	2020-11-16	2020-11-18	Saliva	Positive	Positive
63	M	2020-11-18	2020-11-19	Saliva	Positive	Negative
21	F	2020-11-21	2020-11-22	Saliva	Positive	Positive
33	F	2020-11-21	2020-11-23	Saliva	Positive	Positive
60	F	2020-11-20	2020-11-24	Saliva	Positive	Positive
31	M	2020-11-20	2020-11-24	Saliva	Positive	Positive
37	F	2020-11-15	2020-11-19	Saliva	Positive	Positive
47	M	2020-11-15	2020-11-19	Saliva	Positive	Negative



570 Hood Road, Unit 23,
 Markham, ON
 T: 905-944-9565
 F: 905-944-0406

Clinical Study
 Rapid Response™ COVID-19 Antigen Rapid Test Cassette
 (Oral Fluid)
 COV-2C25B

38	F	2020-11-28	2020-11-30	Saliva	Positive	Positive
38	M	2020-11-19	2020-11-24	Saliva	Positive	Positive
62	M	2020-11-22	2020-11-24	Saliva	Positive	Positive
47	M	2020-11-22	2020-11-24	Saliva	Positive	Positive
24	F	2020-11-18	2020-11-22	Saliva	Positive	Positive
25	M	2020-11-21	2020-11-24	Saliva	Positive	Positive
62	M	2020-11-15	2020-11-18	Saliva	Positive	Positive
49	F	2020-11-19	2020-11-23	Saliva	Positive	Negative
41	M	2020-11-21	2020-11-25	Saliva	Positive	Positive
51	M	2020-11-25	2020-11-29	Saliva	Positive	Positive
47	M	2020-11-14	2020-11-23	Saliva	Positive	Positive
20	F	2020-11-20	2020-11-20	Saliva	Positive	Negative
29	M	2020-11-18	2020-11-19	Saliva	Positive	Positive
53	M	2020-11-30	2020-12-02	Saliva	Positive	Positive
52	M	2020-11-14	2020-11-20	Saliva	Positive	Positive
45	F	2020-11-16	2020-11-23	Saliva	Positive	Positive
31	F	2020-11-24	2020-11-24	Saliva	Positive	Negative
28	M	2020-11-30	2020-12-02	Saliva	Positive	Positive
27	M	2020-11-19	2020-11-23	Saliva	Positive	Positive
21	F	2020-11-18	2020-11-24	Saliva	Positive	Positive
31	M	2020-11-16	2020-11-19	Saliva	Positive	Positive
38	M	2020-11-20	2020-11-24	Saliva	Positive	Positive
19	M	2020-11-21	2020-11-25	Saliva	Positive	Positive
43	F	2020-11-24	2020-11-25	Saliva	Positive	Positive
24	F	2020-11-22	2020-11-23	Saliva	Positive	Positive
66	M	2020-11-21	2020-11-25	Saliva	Positive	Negative
36	F	2020-11-21	2020-11-23	Saliva	Positive	Positive
52	M	2020-11-15	2020-11-23	Saliva	Positive	Positive
36	M	2020-11-28	2020-12-02	Saliva	Positive	Positive
24	M	2020-11-13	2020-11-17	Saliva	Positive	Positive
25	M	2020-11-22	2020-11-25	Saliva	Positive	Positive
41	M	2020-11-28	2020-11-30	Saliva	Positive	Positive
53	M	2020-11-30	2020-12-02	Saliva	Positive	Positive
22	M	2020-11-11	2020-11-19	Saliva	Positive	Positive
42	F	2020-11-24	2020-11-25	Saliva	Positive	Positive
32	F	2020-11-17	2020-11-20	Saliva	Positive	Positive
22	M	2020-11-22	2020-11-24	Saliva	Positive	Positive
27	F	2020-11-19	2020-11-23	Saliva	Positive	Positive
40	F	2020-11-20	2020-11-23	Saliva	Positive	Positive
35	F	2020-11-20	2020-11-23	Saliva	Positive	Positive
21	M	2020-11-25	2020-11-29	Saliva	Positive	Positive
41	M	2020-11-18	2020-11-23	Saliva	Positive	Positive
51	F	2020-11-22	2020-11-24	Saliva	Positive	Positive
23	F	2020-11-17	2020-11-24	Saliva	Positive	Positive
60	M	2020-11-20	2020-11-24	Saliva	Positive	Positive



570 Hood Road, Unit 23,
 Markham, ON
 T: 905-944-9565
 F: 905-944-0406

Clinical Study
 Rapid Response™ COVID-19 Antigen Rapid Test Cassette
 (Oral Fluid)
 COV-2C25B

43	M	2020-11-21	2020-11-23	Saliva	Positive	Positive
42	M	2020-11-21	2020-11-25	Saliva	Positive	Positive
37	M	2020-11-15	2020-11-20	Saliva	Positive	Positive
46	F	2020-11-16	2020-11-23	Saliva	Positive	Positive
61	F	2020-11-18	2020-11-25	Saliva	Positive	Positive
22	F	2020-11-28	2020-11-30	Saliva	Positive	Positive
51	M	2020-11-09	2020-11-19	Saliva	Positive	Positive
25	M	2020-11-17	2020-11-23	Saliva	Positive	Positive
19	M	2020-11-22	2020-11-25	Saliva	Positive	Positive
44	M	2020-11-23	2020-11-24	Saliva	Positive	Positive
15	M	2020-11-16	2020-11-19	Saliva	Positive	Positive
27	M	2020-11-28	2020-11-30	Saliva	Positive	Positive
41	M	2020-11-21	2020-11-25	Saliva	Positive	Negative
39	M	2020-11-28	2020-11-28	Saliva	Positive	Positive
23	M	2020-11-21	2020-11-24	Saliva	Positive	Positive
33	M	2020-11-15	2020-11-19	Saliva	Positive	Positive
32	F	2020-11-12	2020-11-20	Saliva	Positive	Positive
22	F	2020-11-17	2020-11-20	Saliva	Positive	Positive
34	F	2020-11-14	2020-11-20	Saliva	Positive	Negative
18	F	2020-11-29	2020-12-01	Saliva	Positive	Positive
27	M	2020-11-27	2020-11-29	Saliva	Positive	Positive
21	F	2020-11-19	2020-11-22	Saliva	Positive	Positive
40	M	2020-11-10	2020-11-20	Saliva	Positive	Positive
29	F	2020-11-18	2020-11-19	Saliva	Positive	Positive
23	F	2020-11-24	2020-11-29	Saliva	Positive	Negative
50	M	2020-11-14	2020-11-19	Saliva	Positive	Positive
49	F	2020-11-18	2020-11-20	Saliva	Positive	Positive
19	M	2020-11-20	2020-11-25	Saliva	Positive	Positive
18	F	2020-11-18	2020-11-25	Saliva	Positive	Positive
51	F	unknown	2020-12-01	Saliva	Positive	Positive
27	M	2020-11-20	2020-11-24	Saliva	Positive	Positive
58	M	2020-11-22	2020-11-29	Saliva	Positive	Positive
52	M	2020-11-28	2020-12-01	Saliva	Positive	Positive
55	M	2020-11-22	2020-11-24	Saliva	Positive	Positive
25	M	2020-11-20	2020-11-29	Saliva	Positive	Positive
14	M	2020-11-27	2020-11-30	Saliva	Positive	Positive
30	F	2020-11-15	2020-11-19	Saliva	Positive	Positive
47	M	2020-11-19	2020-11-25	Saliva	Positive	Positive
18	M	2020-11-20	2020-11-25	Saliva	Positive	Positive
29	M	2020-11-26	2020-12-02	Saliva	Positive	Positive
27	M	2020-11-01	2020-11-25	Saliva	Positive	Positive
55	F	2020-11-19	2020-11-20	Saliva	Positive	Positive
30	F	2020-11-20	2020-11-23	Saliva	Positive	Positive
24	F	2020-11-28	2020-12-01	Saliva	Positive	Positive